

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
 Data File : VN074422.D
 Acq On : 12 Sep 2022 20:51
 Operator : JC\MD
 Sample : N4558-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-203-205

Quant Time: Sep 13 06:29:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	291447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	495511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	480955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.621	152	215727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	217526	43.139	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.280%
35) Dibromofluoromethane	7.957	113	179013	53.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.020%
50) Toluene-d8	10.380	98	602153	46.197	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.400%
62) 4-Bromofluorobenzene	12.674	95	232588	49.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.300%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.540	73	57582	3.938	ug/l	93
27) cis-1,2-Dichloroethene	7.245	96	1573	0.314	ug/l	88
44) Trichloroethene	9.151	130	35450	9.371	ug/l	90
64) Tetrachloroethene	10.922	164	4196	1.287	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091222\
Data File : VN074422.D
Acq On : 12 Sep 2022 20:51
Operator : JC\MD
Sample : N4558-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-203-205

Quant Time: Sep 13 06:29:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

