

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091223\
 Data File : VN079237.D
 Acq On : 12 Sep 2023 15:53
 Operator : JC\MD
 Sample : 04352-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-02-090723

Quant Time: Sep 13 01:33:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	86732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	141521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	137928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	41176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	58809	50.230	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.460%	
35) Dibromofluoromethane	8.171	113	49200	52.264	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.520%	
50) Toluene-d8	10.571	98	173412	50.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	12.853	95	54098	49.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.140%	
Target Compounds						
					Qvalue	
30) Chloroform	7.971	83	3860	1.857	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091223\
Data File : VN079237.D
Acq On : 12 Sep 2023 15:53
Operator : JC\MD
Sample : 04352-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-02-090723

Quant Time: Sep 13 01:33:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

