

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091223\
 Data File : VN079239.D
 Acq On : 12 Sep 2023 16:47
 Operator : JC\MD
 Sample : 04353-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-090723

Quant Time: Sep 13 01:34:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	85409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	137295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	130830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	40829	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	58174	50.457	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.920%	
35) Dibromofluoromethane	8.171	113	50270	55.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.080%	
50) Toluene-d8	10.571	98	174185	51.944	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.880%	
62) 4-Bromofluorobenzene	12.853	95	53899	50.389	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	2167	4.426	ug/l	99
20) Methylene Chloride	5.277	84	2589	2.173	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091223\
Data File : VN079239.D
Acq On : 12 Sep 2023 16:47
Operator : JC\MD
Sample : 04353-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-090723

Quant Time: Sep 13 01:34:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

