

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
 Data File : VN083835.D
 Acq On : 12 Sep 2024 16:41
 Operator : JC\MD
 Sample : P3905-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: Sep 13 00:20:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	150205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	98674	51.506	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.020%
35) Dibromofluoromethane	8.165	113	78210	46.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	10.565	98	296338	51.115	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.847	95	140947	58.927	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	117.860%
Target Compounds						
16) Acetone	4.430	43	40187	52.328	ug/l	Qvalue 97
20) Methylene Chloride	5.265	84	6532	4.156	ug/l	97
37) Ethyl Acetate	7.559	43	3577	1.561	ug/l #	86
43) Isopropyl Acetate	8.688	43	222849	53.674	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
Data File : VN083835.D
Acq On : 12 Sep 2024 16:41
Operator : JC\MD
Sample : P3905-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: Sep 13 00:20:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

