

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
 Data File : VN083840.D
 Acq On : 12 Sep 2024 18:42
 Operator : JC\MD
 Sample : P3947-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW-1

Quant Time: Sep 13 00:23:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/13/2024
 Supervised By : Mahesh Dadoda 09/13/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	95876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	214159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97789	67.642	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 135.280%#			
35) Dibromofluoromethane	8.165	113	77347	56.416	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 112.840%			
50) Toluene-d8	10.565	98	288064	60.777	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 121.560%#			
62) 4-Bromofluorobenzene	12.847	95	139488	71.332	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 142.660%#			
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	16468	28.416	ug/l #	77
18) Methyl Acetate	5.024	43	7866	4.838	ug/l	94
20) Methylene Chloride	5.259	84	4367	3.682	ug/l #	77
43) Isopropyl Acetate	8.688	43	214061m	63.064	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
Data File : VN083840.D
Acq On : 12 Sep 2024 18:42
Operator : JC\MD
Sample : P3947-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW-1

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/13/2024
Supervised By :Mahesh Dadoda 09/13/2024

Quant Time: Sep 13 00:23:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

