

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
 Data File : VN083849.D
 Acq On : 13 Sep 2024 03:25
 Operator : JC\MD
 Sample : P3845-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PT-VOA-WP

Quant Time: Sep 13 05:58:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	145671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	260331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	226122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109784	49.981	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.960%		
35) Dibromofluoromethane	8.159	113	77597	46.560	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.120%		
50) Toluene-d8	10.565	98	282714	49.069	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.140%		
62) 4-Bromofluorobenzene	12.847	95	110386	46.438	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.880%		
Target Compounds						Qvalue
4) Vinyl Chloride	2.512	62	64154	34.597	ug/l	96
5) Bromomethane	2.901	94	48544	45.596	ug/l	94
16) Acetone	4.430	43	35741	40.590	ug/l	94
19) Methyl tert-butyl Ether	5.794	73	438139	76.847	ug/l	100
21) trans-1,2-Dichloroethene	5.783	96	172613	102.226	ug/l	92
25) 2-Butanone	7.483	43	102372	88.456	ug/l	93
27) cis-1,2-Dichloroethene	7.483	96	42856	21.380	ug/l	90
30) Chloroform	7.959	83	78690	22.261	ug/l	97
40) Benzene	8.606	78	299234	39.632	ug/l	100
42) 1,2-Dichloroethane	8.671	62	393407	139.444	ug/l	100
44) Trichloroethene	9.347	130	95893	54.845	ug/l	95
45) 1,2-Dichloropropane	9.624	63	37653	20.634	ug/l	92
46) Dibromomethane	9.706	93	19254	15.064	ug/l	99
47) Bromodichloromethane	9.882	83	255916	92.514	ug/l	98
52) Toluene	10.629	92	443534	96.039	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	277729	99.050	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	301353	102.316	ug/l	97
60) Dibromochloromethane	11.359	129	58816	30.482	ug/l	100
61) 1,2-Dibromoethane	11.471	107	70288	40.877	ug/l	98
65) Chlorobenzene	11.888	112	226653	45.099	ug/l	99
67) Ethyl Benzene	11.965	91	746385	83.603	ug/l	100
68) m/p-Xylenes	12.070	106	85510	25.798	ug/l	100
69) o-Xylene	12.400	106	341069	106.294	ug/l	99
70) Styrene	12.412	104	539346	102.420	ug/l	98
71) Bromoform	12.576	173	89460	76.198	ug/l #	99
76) 1,2,3-Trichloropropane	12.994	75	213477	95.887	ug/l #	1
80) 1,3,5-Trimethylbenzene	13.170	105	198175	27.536	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	103661	14.339	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	350320	90.339	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	149861	38.166	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	49712	13.283	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	60978	121.963	ug/l	99
93) 1,2,4-Trichlorobenzene	15.388	180	2874	1.431	ug/l	97
95) Naphthalene	15.641	128	457351	74.853	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	2823	1.471	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
Data File : VN083849.D
Acq On : 13 Sep 2024 03:25
Operator : JC\MD
Sample : P3845-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PT-VOA-WP

Quant Time: Sep 13 05:58:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
Data File : VN083849.D
Acq On : 13 Sep 2024 03:25
Operator : JC\MD
Sample : P3845-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PT-VOA-WP

Quant Time: Sep 13 05:58:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

