

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
 Data File : VN083860.D
 Acq On : 13 Sep 2024 07:48
 Operator : JC\MD
 Sample : P3957-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D2-20240911

Quant Time: Sep 13 08:52:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	114956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	240578	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	94909	54.754	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.500%
35) Dibromofluoromethane	8.165	113	74957	48.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	10.565	98	275567	51.756	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.520%
62) 4-Bromofluorobenzene	12.847	95	127275	57.939	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	115.880%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.336	96	2072	1.666	ug/l	# 84
27) cis-1,2-Dichloroethene	7.483	96	81086	51.261	ug/l	92
30) Chloroform	7.959	83	5699	2.043	ug/l	90
42) 1,2-Dichloroethane	8.671	62	4907	1.882	ug/l	98
44) Trichloroethene	9.353	130	648776	401.529	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091224\
Data File : VN083860.D
Acq On : 13 Sep 2024 07:48
Operator : JC\MD
Sample : P3957-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D2-20240911

Quant Time: Sep 13 08:52:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

