


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091318\
Data File : VN051156.D
Acq On : 13 Sep 2018 7:57
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Title : SW846 8260
Last Update : Tue Sep 11 02:27:09 2018



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	9509	PASS
75	95	30	60	44.6	27882	PASS
95	95	100	100	100.0	62474	PASS
96	95	5	9	6.9	4335	PASS
173	174	0.00	2	1.5	861	PASS
174	95	50	100	90.6	56613	PASS
175	174	5	9	7.2	4088	PASS
176	174	95	101	97.4	55125	PASS
177	176	5	9	6.3	3449	PASS