

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057957.D
 Acq On : 12 Sep 2019 10:07
 Operator : JC/SP
 Sample : PB122974TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB122974TB

Quant Time: Sep 13 05:03:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	222933	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	417113	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	369824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	168632	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	332614	73.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	147.32%	
35) Dibromofluoromethane	7.57	113	220762	57.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.68%	
50) Toluene-d8	10.09	98	737583	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	12.40	95	234906	41.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.48%	
Target Compounds						
16) Acetone	3.81	43	15094	10.197	ug/l	Qvalue 100
18) Methyl Acetate	4.31	43	8940	2.265	ug/l #	89
43) Isopropyl Acetate	8.15	43	91423	10.877	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057957.D
Acq On : 12 Sep 2019 10:07
Operator : JC/SP
Sample : PB122974TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB122974TB

Quant Time: Sep 13 05:03:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

