

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057959.D
 Acq On : 12 Sep 2019 10:52
 Operator : JC/SP
 Sample : PB123001TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123001TB

Quant Time: Sep 13 05:06:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	281520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	468482	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	439965	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	230612	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	315396	55.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.62%	
35) Dibromofluoromethane	7.57	113	218915	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
50) Toluene-d8	10.09	98	860633	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.40	95	288565	45.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.22%	
Target Compounds						
16) Acetone	3.81	43	11229	6.007	ug/l	99
43) Isopropyl Acetate	8.16	43	94273	9.987	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057959.D
Acq On : 12 Sep 2019 10:52
Operator : JC/SP
Sample : PB123001TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123001TB

Quant Time: Sep 13 05:06:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

