

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057964.D
 Acq On : 12 Sep 2019 12:42
 Operator : JC/SP
 Sample : K4790-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-3

Quant Time: Sep 13 05:17:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	270515	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	458122	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	428680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	216688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	308180	56.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.50%	
35) Dibromofluoromethane	7.58	113	209501	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.09	98	843286	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	12.41	95	283087	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
Target Compounds						
16) Acetone	3.80	43	16649	9.269	ug/l	96
18) Methyl Acetate	4.32	43	14031	2.930	ug/l #	89
20) Methylene Chloride	4.53	84	13604	2.817	ug/l	93
43) Isopropyl Acetate	8.16	43	89498	9.695	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057964.D
Acq On : 12 Sep 2019 12:42
Operator : JC/SP
Sample : K4790-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-3

Quant Time: Sep 13 05:17:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

