

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057967.D
 Acq On : 12 Sep 2019 13:48
 Operator : JC/SP
 Sample : K4644-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-D2-(4)

Quant Time: Sep 13 05:36:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	287090	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	483987	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	448117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	231476	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	336967	57.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.90%	
35) Dibromofluoromethane	7.58	113	228463	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
50) Toluene-d8	10.09	98	894577	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	12.41	95	298614	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
16) Acetone	3.80	43	14133	7.414	ug/l	95
18) Methyl Acetate	4.31	43	8678	1.707	ug/l #	84
43) Isopropyl Acetate	8.16	43	30911	3.170	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057967.D
Acq On : 12 Sep 2019 13:48
Operator : JC/SP
Sample : K4644-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-D2-(4)

Quant Time: Sep 13 05:36:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

