

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057970.D
 Acq On : 12 Sep 2019 14:52
 Operator : JC/SP
 Sample : K4644-25
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-E1-(0)

Quant Time: Sep 13 05:41:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	279284	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	478533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	439219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	225330	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	318965	56.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.78%	
35) Dibromofluoromethane	7.58	113	219064	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	10.09	98	856829	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	12.41	95	289202	44.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.52%	
Target Compounds						
16) Acetone	3.81	43	14109	7.609	ug/l	97
18) Methyl Acetate	4.31	43	6177	1.249	ug/l #	64
43) Isopropyl Acetate	8.16	43	12971	1.345	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057970.D
Acq On : 12 Sep 2019 14:52
Operator : JC/SP
Sample : K4644-25
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-E1-(0)

Quant Time: Sep 13 05:41:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

