

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057974.D
 Acq On : 12 Sep 2019 17:30
 Operator : JC/SP
 Sample : VN0913WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0913WBL02

Quant Time: Sep 13 05:49:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	290002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	486678	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	438694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	220056	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	316474	53.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.76%	
35) Dibromofluoromethane	7.57	113	216312	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
50) Toluene-d8	10.09	98	859387	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
62) 4-Bromofluorobenzene	12.41	95	276274	41.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057974.D
Acq On : 12 Sep 2019 17:30
Operator : JC/SP
Sample : VN0913WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0913WBL02

Quant Time: Sep 13 05:49:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

