

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057979.D
 Acq On : 12 Sep 2019 19:21
 Operator : JC/SP
 Sample : K4859-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IDW-GW807-091119

Quant Time: Sep 13 06:28:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	256418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	434215	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	390023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	211864	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	299086	57.59	ug/l	0.00
Spiked Amount			Recovery	=	115.18%	
35) Dibromofluoromethane	7.58	113	212672	53.07	ug/l	0.00
Spiked Amount			Recovery	=	106.14%	
50) Toluene-d8	10.09	98	782050	50.26	ug/l	0.00
Spiked Amount			Recovery	=	100.52%	
62) 4-Bromofluorobenzene	12.41	95	271239	45.74	ug/l	0.00
Spiked Amount			Recovery	=	91.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.80	43	133239	78.259	ug/l	100
18) Methyl Acetate	4.32	43	5949	1.311	ug/l #	76
21) trans-1,2-Dichloroethene	5.03	96	4264	1.445	ug/l #	85
25) 2-Butanone	6.82	43	21865	8.873	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	19923	5.037	ug/l	90
30) Chloroform	7.36	83	96502	13.477	ug/l	98
40) Benzene	8.03	78	7499	0.495	ug/l	96
44) Trichloroethene	8.83	130	56155	15.101	ug/l	95
52) Toluene	10.16	92	8818	0.923	ug/l	98
64) Tetrachloroethene	10.63	164	16712	5.803	ug/l	93
67) Ethyl Benzene	11.51	91	8914	0.498	ug/l	98
68) m/p-Xylenes	11.62	106	5845	0.892	ug/l	92
69) o-Xylene	11.95	106	13182	2.024	ug/l	94
95) Naphthalene	15.13	128	52803	2.773	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057979.D
Acq On : 12 Sep 2019 19:21
Operator : JC/SP
Sample : K4859-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IDW-GW807-091119

Quant Time: Sep 13 06:28:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

