

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057986.D
 Acq On : 12 Sep 2019 21:56
 Operator : JC/SP
 Sample : K4680-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-3-(1.5)

Manual Integrations
 APPROVED

MMDadoda
 9/13/2019 1:04:02 PM

Quant Time: Sep 13 06:41:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	239224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	428729	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	381773	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	188232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	321349	66.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	132.64%	
35) Dibromofluoromethane	7.58	113	220182	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
50) Toluene-d8	10.09	98	776871	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
62) 4-Bromofluorobenzene	12.41	95	252276	43.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.18%	
Target Compounds						
16) Acetone	3.80	43	9657	6.080	ug/l #	88
18) Methyl Acetate	4.31	43	7229m	1.707	ug/l	
43) Isopropyl Acetate	8.16	43	112045	12.970	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057986.D
Acq On : 12 Sep 2019 21:56
Operator : JC/SP
Sample : K4680-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-3-(1.5)

Manual Integrations
APPROVED

MMDadoda
9/13/2019 1:04:02 PM

Quant Time: Sep 13 06:41:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

