

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057989.D
 Acq On : 12 Sep 2019 23:02
 Operator : JC/SP
 Sample : K4680-28
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T2-3-(2)

Quant Time: Sep 13 06:44:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	256565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	466692	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	414229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	207977	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	325802	62.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.40%	
35) Dibromofluoromethane	7.58	113	223033	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
50) Toluene-d8	10.09	98	819469	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	12.41	95	270170	42.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.78%	
Target Compounds						
16) Acetone	3.81	43	9667	5.675	ug/l #	82
18) Methyl Acetate	4.31	43	6177	1.360	ug/l #	74
43) Isopropyl Acetate	8.16	43	144717	15.389	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057989.D
Acq On : 12 Sep 2019 23:02
Operator : JC/SP
Sample : K4680-28
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T2-3-(2)

Quant Time: Sep 13 06:44:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

