

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091319\
 Data File : VN057997.D
 Acq On : 13 Sep 2019 1:59
 Operator : JC/SP
 Sample : K4644-27
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-E3-(0)

Quant Time: Sep 13 06:54:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	271562	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	476681	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	437196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	228867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	324038	58.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.82%	
35) Dibromofluoromethane	7.58	113	226937	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	10.09	98	863224	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
62) 4-Bromofluorobenzene	12.41	95	288027	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
Target Compounds						
16) Acetone	3.80	43	14672	8.137	ug/l	97
18) Methyl Acetate	4.31	43	48621	10.113	ug/l	99
43) Isopropyl Acetate	8.16	43	40861	4.254	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091319\
Data File : VN057997.D
Acq On : 13 Sep 2019 1:59
Operator : JC/SP
Sample : K4644-27
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-E3-(0)

Quant Time: Sep 13 06:54:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

