

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091321\
 Data File : VN068461.D
 Acq On : 13 Sep 2021 16:48
 Operator : JC/MD
 Sample : VN0913WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0913WBL01

Quant Time: Sep 14 04:27:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N091321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Sep 14 04:23:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	96618	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	503897	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	436432	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	182951	32.063	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.867%
60) 4-Bromofluorobenzene	12.734	95	163439	26.274	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.567%
63) Toluene-d8	10.446	98	583858	30.752	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.500%
Target Compounds						Qvalue
18) Methylene Chloride	5.095	84	17324	3.287	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091321\
Data File : VN068461.D
Acq On : 13 Sep 2021 16:48
Operator : JC/MD
Sample : VN0913WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0913WBL01

Quant Time: Sep 14 04:27:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N091321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Sep 14 04:23:31 2021
Response via : Initial Calibration

