

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074425.D
 Acq On : 13 Sep 2022 11:43
 Operator : JC\MD
 Sample : VN0913MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0913MBL01

Quant Time: Sep 13 17:13:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	202733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	386474	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	370864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	161538	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	186727	53.235	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.480%
35) Dibromofluoromethane	7.951	113	144121	55.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.480%
50) Toluene-d8	10.380	98	489925	48.192	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.380%
62) 4-Bromofluorobenzene	12.674	95	181127	49.073	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074425.D
Acq On : 13 Sep 2022 11:43
Operator : JC\MD
Sample : VN0913MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0913MBL01

Quant Time: Sep 13 17:13:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

