

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074428.D
 Acq On : 13 Sep 2022 13:10
 Operator : JC\MD
 Sample : N4558-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-218-220

Quant Time: Sep 13 17:14:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	206281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	382568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	379355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	167019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	171942	48.177	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.360%
35) Dibromofluoromethane	7.951	113	138913	53.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.580%
50) Toluene-d8	10.380	98	469078	46.612	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.220%
62) 4-Bromofluorobenzene	12.674	95	186876	51.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	5.528	73	11756	1.136	ug/l #	83
44) Trichloroethene	9.151	130	67349	23.060	ug/l	89
64) Tetrachloroethene	10.916	164	4747	1.846	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074428.D
 Acq On : 13 Sep 2022 13:10
 Operator : JC\MD
 Sample : N4558-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-218-220

Quant Time: Sep 13 17:14:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

