

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074430.D
 Acq On : 13 Sep 2022 13:59
 Operator : JC\MD
 Sample : N4471-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-083122

Quant Time: Sep 13 17:15:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	207250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	394453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	380506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	161457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	245729	68.530	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 137.060%#	
35) Dibromofluoromethane	7.951	113	193070	72.502	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 145.000%#	
50) Toluene-d8	10.381	98	656210	63.243	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 126.480%#	
62) 4-Bromofluorobenzene	12.675	95	231293	61.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 122.800%	
Target Compounds						
16) Acetone	4.228	43	22072	11.402	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074430.D
Acq On : 13 Sep 2022 13:59
Operator : JC\MD
Sample : N4471-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-083122

Quant Time: Sep 13 17:15:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

