

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074433.D
 Acq On : 13 Sep 2022 15:13
 Operator : JC\MD
 Sample : N4580-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CO-1

Quant Time: Sep 13 17:17:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	193642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	352773	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	347101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	154949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	164963	49.238	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.480%
35) Dibromofluoromethane	7.957	113	130341	54.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.460%
50) Toluene-d8	10.380	98	445829	48.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.080%
62) 4-Bromofluorobenzene	12.674	95	170014	50.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.920%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.293	59	23598	26.929	ug/l #	90
16) Acetone	4.222	43	40253	22.255	ug/l	99
25) 2-Butanone	7.269	43	10434	3.437	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074433.D
Acq On : 13 Sep 2022 15:13
Operator : JC\MD
Sample : N4580-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CO-1

Quant Time: Sep 13 17:17:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

