

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074434.D
 Acq On : 13 Sep 2022 15:37
 Operator : JC\MD
 Sample : N4588-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1666

Quant Time: Sep 13 17:17:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	195615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	371372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	366844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.621	152	161373	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	171167	50.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.140%
35) Dibromofluoromethane	7.951	113	139294	55.559	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.120%
50) Toluene-d8	10.380	98	468533	47.962	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.920%
62) 4-Bromofluorobenzene	12.674	95	182553	51.471	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074434.D
Acq On : 13 Sep 2022 15:37
Operator : JC\MD
Sample : N4588-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1666

Quant Time: Sep 13 17:17:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

