

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074438.D
 Acq On : 13 Sep 2022 17:16
 Operator : JC\MD
 Sample : N4492-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-01-1.5-11.5-090122

Quant Time: Sep 14 08:03:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	205866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.910	114	384339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	371008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.621	152	175965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	178976	50.249	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.500%	
35) Dibromofluoromethane	7.951	113	140644	54.205	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.420%	
50) Toluene-d8	10.380	98	477306	47.211	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.420%	
62) 4-Bromofluorobenzene	12.674	95	193745	52.783	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.560%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.134	101	8749	3.317	ug/l	97
27) cis-1,2-Dichloroethene	7.251	96	74495	21.027	ug/l	91
44) Trichloroethene	9.157	130	81085	27.635	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074438.D
Acq On : 13 Sep 2022 17:16
Operator : JC\MD
Sample : N4492-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-01-1.5-11.5-090122

Quant Time: Sep 14 08:03:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

