

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074440.D
 Acq On : 13 Sep 2022 18:06
 Operator : JC\MD
 Sample : N4492-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14B-43-48-090122

Quant Time: Sep 14 08:03:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	197878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	371309	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	363964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	165195	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	170324	49.750	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.500%
35) Dibromofluoromethane	7.951	113	134059	53.480	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.960%
50) Toluene-d8	10.381	98	463022	47.406	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.820%
62) 4-Bromofluorobenzene	12.675	95	177255	49.985	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074440.D
Acq On : 13 Sep 2022 18:06
Operator : JC\MD
Sample : N4492-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14B-43-48-090122

Quant Time: Sep 14 08:03:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

