

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074443.D
 Acq On : 13 Sep 2022 19:20
 Operator : JC\MD
 Sample : N4492-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-090122

Quant Time: Sep 14 08:04:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	205397	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	382341	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	370312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.621	152	168588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	176761	49.740	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.480%
35) Dibromofluoromethane	7.951	113	141465	54.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.620%
50) Toluene-d8	10.380	98	472392	46.969	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.940%
62) 4-Bromofluorobenzene	12.674	95	184989	50.661	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.320%
Target Compounds						
					Qvalue	
16) Acetone	4.216	43	20396	10.631	ug/l #	89
25) 2-Butanone	7.263	43	7463	2.317	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074443.D
Acq On : 13 Sep 2022 19:20
Operator : JC\MD
Sample : N4492-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-090122

Quant Time: Sep 14 08:04:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

