

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
 Data File : VN074447.D
 Acq On : 13 Sep 2022 20:58
 Operator : JC\MD
 Sample : N4492-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BR-01-180-200-090122-H

Quant Time: Sep 14 08:06:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 06:10:29 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/14/2022
 Supervised By : Mahesh Dadoda 09/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	189871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	351347	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	347014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.621	152	156185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	166833	50.786	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.580%			
35) Dibromofluoromethane	7.951	113	130051	54.829	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.660%			
50) Toluene-d8	10.380	98	432684	46.816	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.640%			
62) 4-Bromofluorobenzene	12.674	95	174104	51.886	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.780%			
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	4.122	96	25612	10.821	ug/l	88
16) Acetone	4.234	43	5900	3.327	ug/l #	83
17) Carbon Disulfide	4.463	76	5192m	0.807	ug/l	
24) 1,1-Dichloroethane	6.292	63	7592	1.413	ug/l #	85
27) cis-1,2-Dichloroethene	7.257	96	12410	3.798	ug/l	81
30) Chloroform	7.751	83	17991	3.356	ug/l	92
42) 1,2-Dichloroethane	8.475	62	3369	0.737	ug/l	80
44) Trichloroethene	9.157	130	9256	3.451	ug/l	78
52) Toluene	10.445	92	5614	0.759	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074447.D
Acq On : 13 Sep 2022 20:58
Operator : JC\MD
Sample : N4492-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BR-01-180-200-090122-H

Quant Time: Sep 14 08:06:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 06:10:29 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/14/2022
Supervised By :Mahesh Dadoda 09/14/2022

