


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091322\
Data File : VN074423.D
Acq On : 13 Sep 2022 10:28
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090922W.M
Title : SW846 8260
Last Update : Sat Sep 10 06:10:29 2022



AutoFind: Scans 1833, 1834, 1835; Background Corrected with Scan 1825

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	26.9	10869	PASS
75	95	30	60	59.1	23917	PASS
95	95	100	100	100.0	40472	PASS
96	95	5	9	5.8	2352	PASS
173	174	0.00	2	1.5	438	PASS
174	95	50	100	70.4	28499	PASS
175	174	5	9	8.3	2374	PASS
176	174	95	101	99.4	28317	PASS
177	176	5	9	5.8	1653	PASS

Average of 12.668 to 12.680 min.: VN074423.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.80	131	45.90	98	55.20	154	68.00	552
36.05	541	46.10	178	55.95	1065	68.95	571
36.95	3177	46.90	983	56.95	2496	69.90	35
38.00	3649	47.85	376	59.95	507	70.10	16
39.00	1326	48.15	294	60.95	3161	71.85	524
39.70	106	48.90	1166	62.00	3393	73.00	2611
40.00	13	49.05	1558	63.00	2376	73.95	9597
42.20	70	50.00	10869	63.90	106	75.00	23917
43.90	1015	51.00	3441	64.15	170	76.00	1975
44.80	166	52.00	75	65.20	119	76.80	103
45.00	545	54.90	153	66.80	183	77.10	163

Average of 12.668 to 12.680 min.: VN074423.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.70	65	91.90	1445	117.05	334	142.85	960
78.10	78	93.00	2278	117.70	100	144.10	52
78.85	3159	94.00	6203	118.05	194	147.80	80
79.80	250	95.00	40472	119.05	372	171.50	129
79.95	789	96.05	2352	127.60	89	171.90	105
80.80	2847	103.75	182	127.90	62	172.20	160
81.85	620	104.00	73	128.70	71	172.80	438
86.20	118	105.90	369	129.90	91	173.95	28499
86.90	1836	113.00	92	131.00	137	174.95	2374
87.95	1989	115.90	279	140.75	387	175.95	28317
90.95	410	116.70	311	141.05	520	176.95	1653

Average of 12.668 to 12.680 min.: VN074423.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
177.70	70						
206.90	54						
281.10	120						