

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091418\
 Data File : VN051192.D
 Acq On : 14 Sep 2018 16:01
 Operator : MD\SY
 Sample : PB112829ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112829ZHE#02

Quant Time: Sep 15 00:26:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	546236	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	883540	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	729951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	201280	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	334346	55.00	ug/l	0.00
Spiked Amount			Recovery	=	110.00%	
35) Dibromofluoromethane	7.59	113	335682	50.56	ug/l	0.00
Spiked Amount			Recovery	=	101.12%	
50) Toluene-d8	10.09	98	1186373	47.09	ug/l	0.00
Spiked Amount			Recovery	=	94.18%	
62) 4-Bromofluorobenzene	12.40	95	289086	34.69	ug/l	0.00
Spiked Amount			Recovery	=	69.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	37428	19.52	ug/l	92
18) Methyl Acetate	4.34	43	9893	2.03	ug/l	91
20) Methylene Chloride	4.55	84	6893	1.03	ug/l	90
43) Isopropyl Acetate	8.17	43	258530	26.08	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091418\
Data File : VN051192.D
Acq On : 14 Sep 2018 16:01
Operator : MD\SY
Sample : PB112829ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112829ZHE#02

Quant Time: Sep 15 00:26:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

