

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091418\
 Data File : VN051204.D
 Acq On : 14 Sep 2018 21:44
 Operator : MD\SY
 Sample : PB112829ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112829ZHE#09

Quant Time: Sep 15 01:27:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	526600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	845463	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	691884	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	197804	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	325020	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
35) Dibromofluoromethane	7.59	113	324129	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
50) Toluene-d8	10.09	98	1132333	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
62) 4-Bromofluorobenzene	12.40	95	278147	34.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	36931	20.12	ug/l	96
18) Methyl Acetate	4.34	43	9786	2.09	ug/l #	61
29) Tetrahydrofuran	7.23	42	1569	1.09	ug/l #	50
43) Isopropyl Acetate	8.17	43	229625	24.21	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091418\
Data File : VN051204.D
Acq On : 14 Sep 2018 21:44
Operator : MD\SY
Sample : PB112829ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112829ZHE#09

Quant Time: Sep 15 01:27:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

