

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058007.D
 Acq On : 13 Sep 2019 10:56
 Operator : JC/SP
 Sample : K4823-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

MMDadoda
 9/16/2019 2:46:41 PM

Quant Time: Sep 14 01:54:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N090919W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Sep 09 12:14:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	68379	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	408427	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	343158	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.01	65	192133	31.77	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.90%
60) 4-Bromofluorobenzene	12.40	95	158059	26.92	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	89.73%
63) Toluene-d8	10.09	98	501470	30.45	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.50%

Target Compounds				Ovalue
12) Methyl Acetate	4.32	43	10383m	2.093 ug/l

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091419\
Data File : VN058007.D
Acq On : 13 Sep 2019 10:56
Operator : JC/SP
Sample : K4823-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Manual Integrations
APPROVED

MMDadoda
9/16/2019 2:46:41 PM

Quant Time: Sep 14 01:54:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N090919W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Sep 09 12:14:07 2019
Response via : Initial Calibration

