

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058029.D
 Acq On : 13 Sep 2019 20:07
 Operator : JC/SP
 Sample : K4680-36
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-3-(1.5)

Quant Time: Sep 14 05:35:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	318982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	534874	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	492708	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	252453	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	338581	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	7.57	113	235104	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	10.09	98	898686	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
62) 4-Bromofluorobenzene	12.41	95	313534	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
Target Compounds						
16) Acetone	3.80	43	15248	7.199	ug/l	98
20) Methylene Chloride	4.53	84	9269	1.095	ug/l	90
43) Isopropyl Acetate	8.15	43	98414	9.131	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091419\
Data File : VN058029.D
Acq On : 13 Sep 2019 20:07
Operator : JC/SP
Sample : K4680-36
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
T3-3-(1.5)

Quant Time: Sep 14 05:35:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

