

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058038.D
 Acq On : 13 Sep 2019 23:25
 Operator : JC/SP
 Sample : K4694-28
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-4-(0)

Quant Time: Sep 14 05:58:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	283458	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	506550	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	452115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	222235	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	337245	58.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.48%	
35) Dibromofluoromethane	7.58	113	229553	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	10.08	98	796766	43.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.78%	
62) 4-Bromofluorobenzene	12.41	95	273989	39.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.22%	
Target Compounds						
16) Acetone	3.81	43	10500	5.579	ug/l #	88
43) Isopropyl Acetate	8.15	43	117897	11.551	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091419\
Data File : VN058038.D
Acq On : 13 Sep 2019 23:25
Operator : JC/SP
Sample : K4694-28
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-4-(0)

Quant Time: Sep 14 05:58:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

