

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058051.D
 Acq On : 14 Sep 2019 4:54
 Operator : JC/SP
 Sample : VN0914WBL04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0914WBL04

Quant Time: Sep 14 06:49:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	321498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	533204	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	472555	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	251322	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	348186	53.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.94%	
35) Dibromofluoromethane	7.58	113	240354	48.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.68%	
50) Toluene-d8	10.09	98	941184	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
62) 4-Bromofluorobenzene	12.41	95	310463	42.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

