

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058061.D
 Acq On : 14 Sep 2019 8:55
 Operator : JC/SP
 Sample : K4850-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC-7-(5)

Quant Time: Sep 16 05:58:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	334879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	576611	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	535193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282553	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	378663	55.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.66%	
35) Dibromofluoromethane	7.57	113	258082	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
50) Toluene-d8	10.08	98	1046844	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	12.41	95	345285	43.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.70%	
Target Compounds						
16) Acetone	3.80	43	35889	16.141	ug/l	95
20) Methylene Chloride	4.53	84	16267	2.678	ug/l	98
43) Isopropyl Acetate	8.16	43	147847	12.725	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091419\
Data File : VN058061.D
Acq On : 14 Sep 2019 8:55
Operator : JC/SP
Sample : K4850-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AOC-7(5)

Quant Time: Sep 16 05:58:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

