

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091419\
 Data File : VN058071.D
 Acq On : 14 Sep 2019 12:34
 Operator : JC/SP
 Sample : K4873-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-28

Quant Time: Sep 16 06:01:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	324532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	565834	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	519339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	272625	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	380055	57.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.64%	
35) Dibromofluoromethane	7.58	113	256966	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
50) Toluene-d8	10.09	98	1031485	50.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.74%	
62) 4-Bromofluorobenzene	12.41	95	341073	44.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.28%	
Target Compounds						
16) Acetone	3.80	43	16423	7.622	ug/l	96
18) Methyl Acetate	4.31	43	15794	2.749	ug/l	93
20) Methylene Chloride	4.54	84	55559	12.624	ug/l	97
43) Isopropyl Acetate	8.15	43	132172	11.592	ug/l #	83
52) Toluene	10.15	92	14168	1.139	ug/l	88
95) Naphthalene	15.13	128	233093	13.457	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091419\
Data File : VN058071.D
Acq On : 14 Sep 2019 12:34
Operator : JC/SP
Sample : K4873-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 74 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-28

Quant Time: Sep 16 06:01:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

