


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091419\
 Data File : VN058049.D
 Acq On : 14 Sep 2019 3:27
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N090919W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Sep 09 12:14:07 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.9	24848	PASS
75	95	30	60	56.3	58480	PASS
95	95	100	100	100.0	103954	PASS
96	95	5	9	6.8	7050	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.3	79314	PASS
175	174	5	9	7.7	6138	PASS
176	174	95	101	97.2	77077	PASS
177	176	5	9	6.7	5157	PASS