

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091420\
 Data File : VN063332.D
 Acq On : 14 Sep 2020 10:41
 Operator : JC/MD
 Sample : VN0914MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0914MBL01

Quant Time: Sep 15 05:23:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 11 06:02:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	403296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	755920	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	712128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	274243	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	315090	49.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.46%	
35) Dibromofluoromethane	7.54	113	242933	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
50) Toluene-d8	10.06	98	957646	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
62) 4-Bromofluorobenzene	12.38	95	347347	50.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091420\
Data File : VN063332.D
Acq On : 14 Sep 2020 10:41
Operator : JC/MD
Sample : VN0914MBL01
Misc : 5.00µ/10mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0914MBL01

Quant Time: Sep 15 05:23:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 11 06:02:56 2020
Response via : Initial Calibration

