

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091420\
 Data File : VN063352.D
 Acq On : 14 Sep 2020 19:18
 Operator : JC/MD
 Sample : L3988-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-PZ189I-20200909

Quant Time: Sep 15 08:24:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 11 06:02:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	385125	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	707303	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	705608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	276576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	291660	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
35) Dibromofluoromethane	7.55	113	227244	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	10.06	98	913392	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
62) 4-Bromofluorobenzene	12.38	95	338083	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
Target Compounds						
24) 1,1-Dichloroethane	5.81	63	18398	1.993	ug/l	99
64) Tetrachloroethene	10.60	164	1862	0.322	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091420\
Data File : VN063352.D
Acq On : 14 Sep 2020 19:18
Operator : JC/MD
Sample : L3988-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-PZ189I-20200909

Quant Time: Sep 15 08:24:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091020W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 11 06:02:56 2020
Response via : Initial Calibration

