

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083885.D
 Acq On : 14 Sep 2024 13:08
 Operator : JC\MD
 Sample : P3957-03RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW203D-20240909RE

Quant Time: Sep 16 01:33:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257997	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	211657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	82026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	109368	47.520	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.040%
35) Dibromofluoromethane	8.165	113	77919	47.176	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	10.565	98	282007	49.389	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.780%
62) 4-Bromofluorobenzene	12.847	95	91044	38.647	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	77.300%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.489	96	5595	2.664	ug/l	90
44) Trichloroethene	9.347	130	27622	15.941	ug/l	97
64) Tetrachloroethene	11.112	164	1613	1.146	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083885.D
Acq On : 14 Sep 2024 13:08
Operator : JC\MD
Sample : P3957-03RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW203D-20240909RE

Quant Time: Sep 16 01:33:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

