

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083888.D
 Acq On : 14 Sep 2024 14:20
 Operator : JC\MD
 Sample : P3957-12RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW203D1-20240910RE

Quant Time: Sep 16 01:35:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	170045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	246686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124431	48.529	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.060%
35) Dibromofluoromethane	8.165	113	88447	47.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	10.565	98	325290	50.566	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.140%
62) 4-Bromofluorobenzene	12.847	95	108425	40.852	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	81.700%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.347	96	7412	4.030	ug/l	98
19) Methyl tert-butyl Ether	5.800	73	4367	0.656	ug/l #	87
24) 1,1-Dichloroethane	6.571	63	39923	10.272	ug/l	98
44) Trichloroethene	9.353	130	6970	3.570	ug/l	98
64) Tetrachloroethene	11.112	164	2074	1.264	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083888.D
Acq On : 14 Sep 2024 14:20
Operator : JC\MD
Sample : P3957-12RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW203D1-20240910RE

Quant Time: Sep 16 01:35:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

