

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083891.D
 Acq On : 14 Sep 2024 15:32
 Operator : JC\MD
 Sample : P3957-17RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D1-20240911RE

Quant Time: Sep 16 01:36:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115773	50.064	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.120%
35) Dibromofluoromethane	8.165	113	81608	47.218	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.440%
50) Toluene-d8	10.565	98	297853	49.851	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.700%
62) 4-Bromofluorobenzene	12.847	95	102665	41.648	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.300%
Target Compounds						
						Qvalue
44) Trichloroethene	9.353	130	59006	32.543	ug/l	95
64) Tetrachloroethene	11.106	164	3080	2.011	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083891.D
Acq On : 14 Sep 2024 15:32
Operator : JC\MD
Sample : P3957-17RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D1-20240911RE

Quant Time: Sep 16 01:36:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

