

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083892.D
 Acq On : 14 Sep 2024 15:56
 Operator : JC\MD
 Sample : P3957-18RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUPE1-20240911RE

Quant Time: Sep 16 01:36:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	150085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	259771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	212284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84099	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112861	49.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.740%
35) Dibromofluoromethane	8.171	113	79628	47.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	10.565	98	289411	50.340	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.847	95	92698	39.081	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	78.160%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	58691	33.640	ug/l	98
64) Tetrachloroethene	11.106	164	3498	2.478	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083892.D
Acq On : 14 Sep 2024 15:56
Operator : JC\MD
Sample : P3957-18RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUPE1-20240911RE

Quant Time: Sep 16 01:36:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

