

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083893.D
 Acq On : 14 Sep 2024 16:20
 Operator : JC\MD
 Sample : P3886-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-090624

Quant Time: Sep 16 01:37:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	154081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	88274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113341	48.784	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.560%
35) Dibromofluoromethane	8.165	113	81317	47.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.760%
50) Toluene-d8	10.565	98	289891	49.372	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.847	95	94560	39.035	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	78.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083893.D
Acq On : 14 Sep 2024 16:20
Operator : JC\MD
Sample : P3886-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-090624

Quant Time: Sep 16 01:37:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

