

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091424\  
 Data File : VN083894.D  
 Acq On : 14 Sep 2024 16:44  
 Operator : JC\MD  
 Sample : P3886-06  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TB01-090624

Quant Time: Sep 16 01:37:35 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N091024W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 11 15:18:56 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 152680   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100  | 114   | 264242   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 223317   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.788 | 152   | 86271    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.582  | 65    | 112388   | 48.818   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 97.640%  |
| 35) Dibromofluoromethane    | 8.165  | 113   | 79138    | 46.782   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 93.560%  |
| 50) Toluene-d8              | 10.565 | 98    | 287819   | 49.216   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 98.440%  |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 93449    | 38.731   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =     | 77.460%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN091424\  
Data File : VN083894.D  
Acq On : 14 Sep 2024 16:44  
Operator : JC\MD  
Sample : P3886-06  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
TB01-090624

Quant Time: Sep 16 01:37:35 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N091024W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 11 15:18:56 2024  
Response via : Initial Calibration

