

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
 Data File : VN083899.D
 Acq On : 14 Sep 2024 18:44
 Operator : JC\MD
 Sample : P3947-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW-1

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/16/2024
 Supervised By : Mahesh Dadoda 09/16/2024

Quant Time: Sep 16 01:40:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	195393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	341270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	295156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	124732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	130375	44.251	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.500%
35) Dibromofluoromethane	8.171	113	91624	41.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.880%
50) Toluene-d8	10.565	98	347808	46.050	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.100%
62) 4-Bromofluorobenzene	12.847	95	119609	38.384	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	76.760%#
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	16167	13.688	ug/l	96
20) Methylene Chloride	5.295	84	4032m	1.668	ug/l	
43) Isopropyl Acetate	8.688	43	116438	21.527	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091424\
Data File : VN083899.D
Acq On : 14 Sep 2024 18:44
Operator : JC\MD
Sample : P3947-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW-1

Quant Time: Sep 16 01:40:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/16/2024
Supervised By :Mahesh Dadoda 09/16/2024

