

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051254.D
 Acq On : 16 Sep 2018 5:00
 Operator : MD\SY
 Sample : J4935-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 17 04:27:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	436060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	676998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	546716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	174786	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	240084	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
35) Dibromofluoromethane	7.59	113	255239	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
50) Toluene-d8	10.09	98	863794	44.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.50%	
62) 4-Bromofluorobenzene	12.40	95	226583	35.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.96%	

Target Compounds


						Qvalue
3) Chloromethane	2.06	50	2251	Below Cal		97
5) Bromomethane	2.57	94	1663	Below Cal	#	53
10) Methyl Iodide	3.96	142	96	5.65	ug/l #	51
14) Allyl chloride	4.55	41	210832	29.75	ug/l #	27
16) Acetone	3.83	43	10353	2.86	ug/l	99
18) Methyl Acetate	4.34	43	11695	3.01	ug/l	99
20) Methylene Chloride	4.55	84	6155638	1148.36	ug/l	100
25) 2-Butanone	6.83	43	601	0.32	ug/l #	46
28) Bromochloromethane	7.20	49	3808	0.96	ug/l #	99
29) Tetrahydrofuran	7.23	42	865	0.73	ug/l #	57
30) Chloroform	7.37	83	16346	1.84	ug/l	100
31) Cyclohexane	7.67	56	7358	0.70	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	27825	3.74	ug/l #	48
38) Carbon Tetrachloride	7.67	117	37558	5.13	ug/l #	16
43) Isopropyl Acetate	8.17	43	200436	26.39	ug/l #	82
47) Bromodichloromethane	9.41	83	4569	0.64	ug/l	98
48) Methyl methacrylate	9.18	41	4406	1.19	ug/l #	16
51) 4-Methyl-2-Pentanone	9.99	43	931	0.22	ug/l #	70
52) Toluene	10.16	92	5069	0.37	ug/l	94
56) Ethyl methacrylate	10.27	69	102	2.47	ug/l #	1
59) 2-Hexanone	10.77	43	9884	3.47	ug/l #	58
68) m/p-Xylenes	11.62	106	3456	0.40	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.40	83	60	Below Cal	#	1
76) 1,2,3-Trichloropropane	12.40	75	102563	37.48	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	102563	111.44	ug/l #	15
84) 1,2,4-Trimethylbenzene	13.04	105	2984	0.26	ug/l	99
90) Hexachloroethane	13.90	117	397	Below Cal	#	2
95) Naphthalene	15.13	128	19304	7.96	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051254.D
Acq On : 16 Sep 2018 5:00
Operator : MD\SY
Sample : J4935-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Sep 17 04:27:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 436060

Ion Ratio Lower Upper

168 100

99 48.5 38.3 57.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

200000

150000



100000

50000

0

Time-->

7.60 7.70 7.80



#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 249

Delta R.T. 0.01 min

Lab File: VN051254.D


Acq: 16 Sep 2018 5:00

Tgt Ion: 94 Resp: 1663

Ion Ratio Lower Upper

94 100

96 140.3 75.9 113.9#



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

2.57

1200

1000

800

600


400


200

0

Time-->

2.50 2.55 2.60









#10
Methvl Iodide
Concen: 5.65 ug/l
RT: 3.96 min Scan# 679
Delta R.T. 0.01 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
142	100		
127	71.9	31.0	46.4#
141	0.0	11.3	16.9#



#14
Allyl chloride
Concen: 29.75 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.23 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion	Ratio	Lower	Upper
41	100		
39	0.4	48.8	73.2#
76	0.0	30.6	45.8#





#16

Acetone

Concen: 2.86 ug/l

RT: 3.83 min Scan# 639

Delta R.T. 0.01 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 43 Resp: 10353

Ion Ratio Lower Upper

43 100

58 35.3 27.8 41.6



#18

Methyl Acetate

Concen: 3.01 ug/l

RT: 4.34 min Scan# 797

Delta R.T. 0.01 min

Lab File: VN051254.D




Acq: 16 Sep 2018 5:00


Tgt Ion: 43 Resp: 11695

Ion Ratio Lower Upper

43 100

74 25.6 20.7 31.1









#20
Methylene Chloride
Concen: 1148.36 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 84 Resp: 6155638




Ion	Ratio	Lower	Upper
84	100		
49	113.1	90.2	135.2
51	35.5	28.4	42.6
86	64.3	51.8	77.8




#25
2-Butanone
Concen: 0.32 ug/l
RT: 6.83 min Scan# 1573
Delta R.T. -0.01 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 43 Resp: 601

Ion	Ratio	Lower	Upper
43	100		
72	0.0	23.0	34.6#









#28
Bromochloromethane
Concen: 0.96 ug/l
RT: 7.20 min Scan# 1689
Delta R.T. 0.01 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 49 Resp: 3808
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.4
130 88.1 70.7 106.1




Abundance Ion 48.90 (48.60 to 49.60): VN0
Ion 128.80 (128.50 to 129.50): V
Ion 129.80 (129.50 to 130.50): V






#29
Tetrahydrofuran
Concen: 0.73 ug/l
RT: 7.23 min Scan# 1698
Delta R.T. 0.02 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 42 Resp: 865
Ion Ratio Lower Upper
42 100
72 32.7 37.4 56.0#
71 0.0 34.7 52.1#



Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0









#30
Chloroform
Concen: 1.84 ug/l
RT: 7.37 min Scan# 1741
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 83 Resp: 16346
Ion Ratio Lower Upper
83 100
85 65.7 52.6 79.0



#31
Cyclohexane
Concen: 0.70 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 56 Resp: 7358
Ion Ratio Lower Upper
56 100
69 192.7 26.3 39.5#
84 163.2 69.4 104.0#









#33
 1,2-Dichloroethane-d4
 Concen: 49.47 ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion: 65 Resp: 240084
 Ion Ratio Lower Upper
 65 100
 67 54.6 0.0 112.0



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00

Tgt Ion: 114 Resp: 676998
 Ion Ratio Lower Upper
 114 100
 63 18.3 0.0 37.8
 88 15.1 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 50.17 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
113	100		
111	102.3	83.0	124.6
192	23.5	18.7	28.1




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 3.74 ug/l
 RT: 7.67 min Scan# 1834
 Delta R.T. -0.12 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.2	18.6	55.8#
77	0.0	25.2	37.8#



Abundance Ion 74.90 (74.60 to 75.60): V
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): V









#38
Carbon Tetrachloride
Concen: 5.13 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. -0.11 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:117 Resp: 37558
Ion Ratio Lower Upper
117 100
119 5.1 76.8 115.2#
121 0.0 24.6 36.8#



#43
Isopropyl Acetate
Concen: 26.39 ug/l
RT: 8.17 min Scan# 1989
Delta R.T. 0.01 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 43 Resp: 200436
Ion Ratio Lower Upper
43 100
61 18.2 25.0 37.6#
87 12.3 11.8 17.8









#47
Bromodichloromethane
Concen: 0.64 ug/l
RT: 9.41 min Scan# 2374
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 83 Resp: 4569
Ion Ratio Lower Upper
83 100
85 66.7 52.3 78.5
127 10.1 7.5 11.3




Abundance Ion 82.90 (82.60 to 83.60): VNO
Ion 84.90 (84.60 to 85.60): VNO
Ion 126.80 (126.50 to 127.50): VNO






#48
Methyl methacrylate
Concen: 1.19 ug/l
RT: 9.18 min Scan# 2304
Delta R.T. -0.02 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 41 Resp: 4406
Ion Ratio Lower Upper
41 100
69 0.0 77.7 116.5#
39 13.1 41.6 62.4#



Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO









#50
Toluene-d8
Concen: 44.75 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 98 Resp: 863794
Ion Ratio Lower Upper
98 100
100 64.1 52.0 78.0



#51
4-Methyl-2-Pentanone
Concen: 0.22 ug/l
RT: 9.99 min Scan# 2555
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 43 Resp: 931
Ion Ratio Lower Upper
43 100
58 22.8 33.5 50.3#





#52

Toluene

Concen: 0.37 ug/l

RT: 10.16 min Scan# 2607

Delta R.T. 0.00 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 92 Resp: 5069

Ion Ratio Lower Upper

92 100

91 182.3 139.5 209.3



#56

Ethyl methacrylate

Concen: 2.47 ug/l

RT: 10.27 min Scan# 2644

Delta R.T. -0.15 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00




Tgt Ion: 69 Resp: 102


Ion Ratio Lower Upper

69 100

41 721.6 47.8 71.8#

39 302.9 23.0 34.4#









#59
2-Hexanone
Concen: 3.47 ug/l
RT: 10.77 min Scan# 2797
Delta R.T. 0.02 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 9884
Ion Ratio Lower Upper
43 100
58 27.4 29.3 88.0#



#62
4-Bromofluorobenzene
Concen: 35.48 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion: 95 Resp: 226583
Ion Ratio Lower Upper
95 100
174 96.0 0.0 184.8
176 90.6 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:117 Resp: 546716
Ion Ratio Lower Upper
117 100
82 52.5 41.8 62.6
119 32.6 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V






#68
m/p-Xylenes
Concen: 0.40 ug/l
RT: 11.62 min Scan# 3062
Delta R.T. -0.00 min
Lab File: VN051254.D
Acq: 16 Sep 2018 5:00

Tgt Ion:106 Resp: 3456
Ion Ratio Lower Upper
106 100
91 191.7 159.8 239.6



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:152 Resp: 174786

Ion Ratio Lower Upper

152 100

115 54.6 28.2 84.7




150 158.1 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.40 min Scan# 3305

Delta R.T. -0.10 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00


Tgt Ion: 83 Resp: 60

Ion Ratio Lower Upper

83 100

131 506.7 5.3 16.1#



85 0.0 32.3 96.8#




Abundance Ion 82.90 (82.60 to 83.60): V

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): V





#76

1,2,3-Trichloropropane

Concen: 37.48 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 102563

Ion Ratio Lower Upper




75 100

77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO



#81

trans-1,4-Dichloro-2-butene

Concen: 111.44 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00


Tgt Ion: 75 Resp: 102563

Ion Ratio Lower Upper

75 100

53 0.1 67.8 101.8#



89 0.0 36.0 54.0#




Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO

Ion 88.90 (88.60 to 89.60): VNO









#84
 1,2,4-Trimethylbenzene
 Concen: 0.26 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion:105 Resp: 2984
 Ion Ratio Lower Upper
 105 100
 120 46.1 23.3 69.8



#90
 Hexachloroethane
 Concen: Below Cal
 RT: 13.90 min Scan# 3770
 Delta R.T. 0.02 min
 Lab File: VN051254.D
 Acq: 16 Sep 2018 5:00

Tgt Ion:117 Resp: 397
 Ion Ratio Lower Upper
 117 100
 201 0.0 47.3 142.0#





#95

Naphthalene

Concen: 7.96 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN051254.D

Acq: 16 Sep 2018 5:00

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:128 Resp: 19304

Ion Ratio Lower Upper

128 100

127 14.2 10.6 15.8

129 10.5 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V
15000 Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

