

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051218.D
 Acq On : 15 Sep 2018 13:27
 Operator : MD\SY
 Sample : PB112879TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879TB

Quant Time: Sep 17 03:08:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	451129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	674922	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	552707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	158967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	235936	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
35) Dibromofluoromethane	7.59	113	254079	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
50) Toluene-d8	10.09	98	876250	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
62) 4-Bromofluorobenzene	12.40	95	225876	35.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.96%	
Target Compounds						
16) Acetone	3.82	43	8863	1.33	ug/l	91
18) Methyl Acetate	4.33	43	8586	2.14	ug/l #	69
20) Methylene Chloride	4.56	84	6937	1.25	ug/l	85
43) Isopropyl Acetate	8.17	43	204840	27.05	ug/l #	87
95) Naphthalene	15.13	128	6109	6.25	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051218.D
Acq On : 15 Sep 2018 13:27
Operator : MD\SY
Sample : PB112879TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879TB

Quant Time: Sep 17 03:08:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

