

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091518\
 Data File : VN051220.D
 Acq On : 15 Sep 2018 14:16
 Operator : MD\SY
 Sample : PB112879ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112879ZHE#01

Quant Time: Sep 17 03:18:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 02:27:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	427264	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	643236	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	520311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	136640	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	228163	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
35) Dibromofluoromethane	7.59	113	244204	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
50) Toluene-d8	10.09	98	833548	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
62) 4-Bromofluorobenzene	12.40	95	207143	34.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	9380	2.19	ug/l	96
18) Methyl Acetate	4.34	43	9148	2.40	ug/l	98
20) Methylene Chloride	4.56	84	6281	1.20	ug/l	98
43) Isopropyl Acetate	8.17	43	196230	27.19	ug/l #	88
95) Naphthalene	15.13	128	5680	6.33	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091518\
Data File : VN051220.D
Acq On : 15 Sep 2018 14:16
Operator : MD\SY
Sample : PB112879ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112879ZHE#01

Quant Time: Sep 17 03:18:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091018W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 02:27:09 2018
Response via : Initial Calibration

